



Case Study

 <p>College of American Pathologists 325 Waukegan Road, Northfield, Illinois 60093-27 800-323-4040 • www.cap.org <i>Advancing Excellence</i></p>	CAP Number: Kit # Institution: Attention: City / State:	Kit ID: 20133822 Kit Mailed: 08/28/20XX Original Evaluation: 09/28/20XX Revision Evaluation: 10/13/20XX								
EVALUATION REVISION	CG2-B 2006 Coagulation-Comprehensive Survey									
Test Unit of Measure Peer Group Fibrinogen mg/dL DIAG STAGO DIAG STAGO STA-R/EVLU	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
Reagent/Instrument	Specimen CG2-07 CG2-08 CG2-09 CG2-10 CG2-11	Your Result 273 288 314 276 159	Mean 278.2 292.8 332.5 294.9 166.0	S.D. 12.2 13.3 15.4 15.1 7.3	No. of Labs 121 121 123 122 121	S.D.I. -0.4 -0.4 -1.2 -1.3 -1.0	Limits of Acceptability Lower 222 234 265 235 132	Limits of Acceptability Upper 334 352 399 354 200	Your Grade Acceptable Acceptable Acceptable Acceptable Acceptable	
Factor IX % DIAG STAGO STA-PTT A DIAG STAGO STA-R/EVLU	CG2-09 CG2-10 CG2-11	13 110 60	14.4 125.82 73.8	2.1 15.20 11.5	69 68 71	-0.7 -1.0 -1.2			[26] [26] [26]	
Protein S Total,Ag % MICROLATES PART MED IA DIAG STAGO LIATEST	CG2-07 CG2-08 CG2-11								[42] [42] [42]	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.